

**§ 801.9 Tolerances for test weight apparatuses.**

The maintenance tolerances for test weight per bushel apparatuses used in performing official inspection services shall be:

Item	Tolerance
Beam/scale accuracy ....	±0.10 pound per bushel deviation at any reading, using test weights
Overall accuracy .....	±0.15 pound per bushel, mean deviation from standard test weight apparatus using wheat

**§ 801.10 [Reserved]****§ 801.11 Related design requirements.**

(a) *Suitability.* The design, construction, and location of official sampling and inspection equipment and related sample handling systems shall be suitable for the official sampling and inspection activities for which the equipment is to be used.

(b) *Durability.* The design, construction, and material used in official sampling and inspection equipment and related sample handling systems shall assure that, under normal operating conditions, operating parts will remain fully operable, adjustments will remain reasonably constant, and accuracy will be maintained between equipment test periods.

(c) *Marking and identification.* Official sampling and inspection equipment for which tolerances have been established shall be permanently marked to show the manufacturer's name, initials, or trademark; the serial number of the equipment; and the model, the type, and the design or pattern of the equipment. Operational controls for mechanical samplers and related sample handling systems, including but not limited to pushbuttons and switches, shall be conspicuously identified as to the equipment or activity controlled by the pushbutton or switch.

(d) *Repeatability.* Official inspection equipment when tested in accordance with §§ 800.217 and 800.219 shall, within the tolerances prescribed in §§ 801.3 through 801.10, be capable of repeating its results when the equipment is operated in its normal manner.

(e) *Security.* Mechanical samplers and related sample handling systems shall provide a ready means of sealing to

deter unauthorized adjustments, removal, or changing of component parts or timing sequence without removing or breaking the seals; and otherwise be designed, constructed, and installed in a manner to prevent deception by any person.

(f) *Installation requirements.* Official sampling and inspection equipment and related sample handling systems shall be installed (1) at a site approved by the Service, (2) according to the manufacturer's instructions, and (3) in such a manner that neither the operation nor the performance of the equipment or system will be adversely affected by the foundation, supports, or any other characteristic of the installation.

**§ 801.12 Design requirements incorporated by reference.**

(a) *Moisture meters.* All moisture meters approved for use in official grain moisture determination and certification shall meet applicable requirements contained in the FGIS Moisture Handbook and the General Code and Grain Moisture Meters Code of the 1991 edition of the National Institute of Standards and Technology's (NIST) Handbook 44, "Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices." Pursuant to the provisions of 5 U.S.C. 552(a), the materials in Handbook 44 are incorporated by reference as they exist on the date of approval and a notice of any change in these materials will be published in the FEDERAL REGISTER.

The NIST Handbook is for sale by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20403. It is also available for inspection at the National Archives and Records Administration (NARA). For information on the availability of this material at NARA, call 202-741-6030, or go to: [http://www.archives.gov/federal\\_register/code\\_of\\_federal\\_regulations/ibr\\_locations.html](http://www.archives.gov/federal_register/code_of_federal_regulations/ibr_locations.html).

The following Handbook 44 requirements are not incorporated by reference:

*General Code (1.10.)*

G-S.5.5. Money Values, Mathematical Agreement

G-T.1. Acceptance Tolerances